



UNITED STATES PATENT AND TRADEMARK OFFICE

UNITED STATES DEPARTMENT OF COMMERCE
United States Patent and Trademark Office
Address: COMMISSIONER OF PATENTS AND TRADEMARKS
Washington, D.C. 20231
www.uspto.gov

APPLICATION NO.	FILING DATE	FIRST NAMED INVENTOR	ATTORNEY DOCKET NO.	CONFIRMATION NO.
09/364,967	07/31/1999	KEVIN J. KELLY	P-8035	1149

27581 7590 02/12/2003

MEDTRONIC, INC.
710 MEDTRONIC PARKWAY NE
MS-LC340
MINNEAPOLIS, MN 55432-5604

EXAMINER

TSAI, CAROL S W

ART UNIT PAPER NUMBER

2857

DATE MAILED: 02/12/2003

Please find below and/or attached an Office communication concerning this application or proceeding.

Office Action Summary

Application No.

09/364,967

Applicant(s)

KELLY ET AL.

Examiner

Carol S Tsai

Art Unit

2857

-- The MAILING DATE of this communication appears on the cover sheet with the correspondence address --

Period for Reply

A SHORTENED STATUTORY PERIOD FOR REPLY IS SET TO EXPIRE 3 MONTH(S) FROM THE MAILING DATE OF THIS COMMUNICATION.

- Extensions of time may be available under the provisions of 37 CFR 1.136(a). In no event, however, may a reply be timely filed after SIX (6) MONTHS from the mailing date of this communication.
- If the period for reply specified above is less than thirty (30) days, a reply within the statutory minimum of thirty (30) days will be considered timely.
- If NO period for reply is specified above, the maximum statutory period will apply and will expire SIX (6) MONTHS from the mailing date of this communication.
- Failure to reply within the set or extended period for reply will, by statute, cause the application to become ABANDONED (35 U.S.C. § 133).
- Any reply received by the Office later than three months after the mailing date of this communication, even if timely filed, may reduce any earned patent term adjustment. See 37 CFR 1.704(b).

Status

- 1) ☒ Responsive to communication(s) filed on 18 December 2002.
- 2a) ☐ This action is **FINAL**. 2b) ☒ This action is non-final.
- 3) ☐ Since this application is in condition for allowance except for formal matters, prosecution as to the merits is closed in accordance with the practice under *Ex parte Quayle*, 1935 C.D. 11, 453 O.G. 213.

Disposition of Claims

- 4) ☒ Claim(s) 1-39 is/are pending in the application.
- 4a) Of the above claim(s) _____ is/are withdrawn from consideration.
- 5) ☒ Claim(s) 17 and 20 is/are allowed.
- 6) ☒ Claim(s) 1-16, 18, 19 and 21-39 is/are rejected.
- 7) ☐ Claim(s) _____ is/are objected to.
- 8) ☐ Claim(s) _____ are subject to restriction and/or election requirement.

Application Papers

- 9) ☐ The specification is objected to by the Examiner.
- 10) ☐ The drawing(s) filed on _____ is/are: a) ☐ accepted or b) ☐ objected to by the Examiner.
- Applicant may not request that any objection to the drawing(s) be held in abeyance. See 37 CFR 1.85(a).
- 11) ☐ The proposed drawing correction filed on _____ is: a) ☐ approved b) ☐ disapproved by the Examiner.
- If approved, corrected drawings are required in reply to this Office action.
- 12) ☐ The oath or declaration is objected to by the Examiner.

Priority under 35 U.S.C. §§ 119 and 120

- 13) ☐ Acknowledgment is made of a claim for foreign priority under 35 U.S.C. § 119(a)-(d) or (f).
- a) ☐ All b) ☐ Some * c) ☐ None of:
1. ☐ Certified copies of the priority documents have been received.
2. ☐ Certified copies of the priority documents have been received in Application No. _____.
3. ☐ Copies of the certified copies of the priority documents have been received in this National Stage application from the International Bureau (PCT Rule 17.2(a)).
- * See the attached detailed Office action for a list of the certified copies not received.
- 14) ☐ Acknowledgment is made of a claim for domestic priority under 35 U.S.C. § 119(e) (to a provisional application).
- a) ☐ The translation of the foreign language provisional application has been received.
- 15) ☐ Acknowledgment is made of a claim for domestic priority under 35 U.S.C. §§ 120 and/or 121.

Attachment(s)

- 1) ☒ Notice of References Cited (PTO-892) 4) ☐ Interview Summary (PTO-413) Paper No(s). _____
- 2) ☐ Notice of Draftsperson's Patent Drawing Review (PTO-948) 5) ☐ Notice of Informal Patent Application (PTO-152)
- 3) ☐ Information Disclosure Statement(s) (PTO-1449) Paper No(s) _____ 6) ☐ Other: _____

DETAILED ACTION

1. The text of those sections of Title 35, U.S. Code not included in this action can be found in a prior Office action.

Claim Objections

2. Claims 30-36 are objected to because of the following informalities:

In claim 30, line 3, "tissue, stimulator" should read - - tissue stimulator - -.

Appropriate correction is required.

Claim Rejections - 35 USC § 102

3. The following is a quotation of the appropriate paragraphs of 35 U.S.C. 102 that form the basis for the rejections under this section made in this Office action:

A person shall be entitled to a patent unless –

(e) the invention was described in a patent granted on an application for patent by another filed in the United States before the invention thereof by the applicant for patent, or on an international application by another who has fulfilled the requirements of paragraphs (1), (2), and (4) of section 371(c) of this title before the invention thereof by the applicant for patent.

The changes made to 35 U.S.C. 102(e) by the American Inventors Protection Act of 1999 (AIPA) do not apply to the examination of this application as the application being examined was not (1) filed on or after November 29, 2000, or (2) voluntarily published under 35 U.S.C. 122(b). Therefore, this application is examined under 35 U.S.C. 102(e) prior to the amendment by the AIPA (pre-AIPA 35 U.S.C. 102(e)).

4. Claims 1-16, 19, and 29-35 are rejected under 35 U.S.C. 102(e) as being anticipated by U. S. Patent No. 6,108,579 to Snell et al.

With respect to claims 1, 3, 4, 12, and 29, Snell et al. disclose a method of determining the current status and remaining life of a power source in an implantable neurological tissue stimulator comprising the steps of: assessing the power source voltage of the power source in an implantable neurological tissue stimulator (see col. 7, lines 29-57); determining, based on the assessed power source voltage, where the power source is in its power source life cycle (see col. 4, lines 1-17; col. 5, lines 1-57; col. 9, lines 12-28; and col. 10, lines 23-55); obtaining a used capacity of the power source and a time that the power source has been operating, wherein the used capacity and the time are actual measurements (see col. 4, lines 29-44; col. 5, lines 39-57; col. 8, lines 44-62; and col. 10, lines 22-27); and determining the remaining life of the power source based on the used capacity of the power (see col. 5, lines 45-57; col. 8, line 62 to col. 9, line 5; and col. 10, line 41 to col. 11, line 3).

As to claim 2, Snell et al. also disclose the power source voltage being done by connecting the power source to an analog to digital (A/D) converter (see col. 7, lines 34-37).

As to claim 5-8, Snell et al. also disclose determining the probable usage rate of the power source and dividing the determined remaining capacity by the probable usage rate of the power source (see col. 4, lines 37-67; col. 6, lines 38-58; and col. 8, line 6 to col. 9, line 5).

As to claims 9-11, Snell et al. also disclose determining the used capacity of the power source since the last time the implantable neurological tissue stimulator was reprogrammed (see col. 10, lines 22-40 and col. 11, lines 4-16).

As to claims 12, 13, 31, and 32, Snell et al. also disclose correlating, in a “look-up table”, the power source voltage assessed in the step of assessing the power source voltage to a predetermined “power source capacity remaining”/ “power source capacity used” value.

It is, however, considered inherent that Snell et al. correlating, in a “look-up table”, the power source voltage assessed in the step of assessing the power source voltage to a predetermined “power source capacity remaining”/ “power source capacity used” value (see col. 8, lines 44-58), because such correlating is known to be a necessary step in order to make a determination of battery end-of-service in a medical device.

As to claims 14 and 15, Snell et al. also disclose determining the power source capacity used/remaining and then subtracting this value from the total source capacity in where the power source capacity remaining is determined (see col. 8, line 59 to col. 9, line 5).

As to claims 16, 19, 33, and 34, Snell et al. also disclose calculating, using the power source voltage determined in the step of assessing the power source voltage of the power source in an implantable neurological tissue stimulator, the remaining power source capacity by a formula (see col. 9, line 36 to col. 10, line 62).

As to claim 30, Snell et al. also disclose a device for determining the current status and remaining life of a power source in an implantable neurological tissue stimulator, the device comprising: an implantable neurological tissue stimulator (an implantable cardiac stimulating device 10 shown on Fig. 1), the implantable neurological tissue stimulator having: a source of power (battery 70 shown on Fig. 1); a voltage determining system (monitoring circuit 80 shown on Fig. 1) for determining the voltage of the source of power (see col. 7, lines 46-57); a programmer (programmer 110 shown on Fig. 1) for creating and processing information to be sent to and received from the implantable neurological tissue stimulator, the programmer including a processor (controller 150 shown on Fig. 1) and a memory (memory 170 shown on Fig. 1) attached thereto; a system (telemetry circuit 100 shown on Fig. 1) for communicating

Art Unit: 2857

information between the implantable neurological tissue stimulator and the programmer; wherein the voltage determining system for determining the voltage of the source of power passes the determined voltage of the source of power to the system for communication (see Fig. 1); and wherein the system for communication passes the determined voltage of the source of power from the implantable neurological tissue stimulator to the programmer and to the processor (see Fig. 1), and wherein the processor determines, based on the determined voltage of the source of power, where the source of power is in its life cycle (see col. 4, lines 1-17; col. 5, lines 1-57; col. 9, lines 12-28; and col. 10, lines 23-55); obtains a used capacity of the power source and a time that the power source has been operating, wherein the used capacity and the time are actual measurements (see col. 4, lines 29-44; col. 5, lines 39-57; col. 8, lines 44-62; and col. 10, lines 22-27); and determines the remaining life of the power source based on the used capacity of the power source and the time that the power source has been operating (see col. 5, lines 45-57; col. 8, line 62 to col. 9, line 5; and col. 10, line 41 to col. 11, line 3).

As to claim 35, Snell et al. also disclose the power source being a battery (battery 70 shown on Fig. 1).

Claim Rejections - 35 USC § 103

5. The following is a quotation of 35 U.S.C. 103(a) which forms the basis for all obviousness rejections set forth in this Office action:

(a) A patent may not be obtained though the invention is not identically disclosed or described as set forth in section 102 of this title, if the differences between the subject matter sought to be patented and the prior art are such that the subject matter as a whole would have been obvious at the time the invention was made to a person having ordinary skill in the art to which said subject matter pertains. Patentability shall not be negated by the manner in which the invention was made.

6. Claims 18 and 21 are rejected under 35 U.S.C. 103(a) as being unpatentable over Snell et al. in view of U. S. Patent No. 5,744,931 to Arai et al.

As noted above, with respect to claims 18 and 21, Snell et al. disclose the claimed invention, except calculating the remaining power source capacity/power source capacity used by using a non-linear formula.

Arai et al. teach calculating the remaining power source capacity/power source capacity used by using a non-linear formula (see col. 2, lines 10-16 and col. 3, lines 11-23).

It would have been obvious to one having ordinary skill in the art at the time the invention was made to modify Snell et al.'s method to include calculating the remaining power source capacity/power source capacity used by using a non-linear formula, as taught by Arai et al., in order to approximate the actual relationship between measured power source and remaining power source capacity for estimating an actual battery remaining capacity voltage.

7. Claims 22-24 are rejected under 35 U.S.C. 103(a) as being unpatentable over Snell et al. in view of U. S. Patent No. 5,344,431 to Merritt et al.

As noted above, Snell et al. disclose the claimed invention, except for informing the user of where in the power source life the power source is.

Merritt et al. teach informing the user of where in the power source life the power source is (see col. 9, line 66 to col. 10, line 2).

It would have been obvious to one having ordinary skill in the art at the time the invention was made to modify Snell et al.'s method to include informing the user of where in the

Art Unit: 2857

power source life the power source is, as taught by Merritt et al., in order to alert the patient that that the battery is almost exhausted.

As to claims 23 and 24, Snell et al. also disclose displaying a representation of the percentage of power source capacity used/remaining (see col. 8, lines 6-28).

8. Claims 25-28 are rejected under 35 U.S.C. 103(a) as being unpatentable over Snell et al. in view of Merritt et al. as applied to claims 1 and 22 above, and further in view of U. S. Patent No. 5,994,876 to Canny et al.

As noted above, with respect to claims 25-27, Snell et al. in combination with Merritt et al. teach all the features of the claimed invention, but do not disclose determining whether the remaining power source capacity falls within a predetermined limit; alerting the user if the remaining power source capacity falls within a predetermined limit; and alerting the user by trigger an alarm.

Canny et al. teach determining whether the remaining power source capacity falls within a predetermined limit (see col. 8, lines 20-26); alerting the user if the remaining power source capacity falls within a predetermined limit; and alerting the user by trigger an alarm (see col. 8, lines 26-28).

It would have been obvious to one having ordinary skill in the art at the time the invention was made to modify Snell et al. in combination with Merritt et al.'s method to include steps of determining whether the remaining power source capacity falls within a predetermined limit; alerting the user if the remaining power source capacity falls within a predetermined limit;

and alerting the user by trigger an alarm, as taught by Canny et al., in order to recharge the battery (Canny et al. col. 8, line 28).

As to claim 28, Snell et al. in combination with Merritt et al. do not disclose triggering an alarm chosen from the group consisting of audible or visual warnings.

Canny et al. teach triggering an alarm chosen from the group consisting of audible or visual warnings (see col. 8, lines 25-28).

It would have been obvious to one having ordinary skill in the art at the time the invention was made to modify Snell et al. in combination with Merritt et al.'s method to include triggering an alarm chosen from the group consisting of audible or visual warnings, as taught by Canny et al., in order to recharge the battery (Canny et al. col. 8, line 28).

9. Claim 36 is rejected under 35 U.S.C. 103(a) as being unpatentable over Snell et al. in view of U. S. Patent No. 6,099,495 to Kinghorn et al.

As noted above, Snell et al. disclose the claimed invention, except for the power source being a capacitor.

Kinghorn et al. teach the power source being a capacitor (see Abstract, lines 1-4; col. 1, lines 65 to col. 2, line 2; and col. 2, lines 61 to col. 3, line 2).

It would have been obvious to one having ordinary skill in the art at the time the invention was made to modify Snell et al.'s method to include the power source being a capacitor, as taught by Kinghorn et al., in order to power an implantable electrical transducer capable of moving from one position to another for providing treatment for the patient (Kinghorn et al. col. 1, line 67 to col. 2, lines 2).

10. Claim 37 is rejected under 35 U.S.C. 103(a) as being unpatentable over Snell et al. in view of U. S. Patent No. 5,807,397 to Barreras.

Snell et al. disclose implanting in the patient a pulse generator (pulse generation and delivery circuit 60 shown on Fig. 1) having a power source (battery 70 shown on Fig. 1), and a lead (lead system 30 shown on Fig. 1)) connected to the pulse generator and stimulating nervous tissue with electrical pulses generated by the pulse generator and communicated by the lead and determining the status and remaining life of the power source (see col. 7, lines 29-57).

Snell et al. do not disclose controlling the pulse generator within preset limits by patient to adjust stimulation of nervous tissue.

Barreras teaches controlling the pulse generator within preset limits by patient to adjust stimulation of nervous tissue (see col. 18, lines 31-46 and col. 19, lines 18-23).

It would have been obvious to one having ordinary skill in the art at the time the invention was made to modify Snell et al.'s method to include controlling the pulse generator within preset limits by patient to adjust stimulation of nervous tissue, as taught by Barreras, in order that stimulation parameters or control values in the implanted stimulator can be adjusted within preset limits to function effectively.

11. Claim 38 is rejected under 35 U.S.C. 103(a) as being unpatentable over Snell et al. in view of Merritt et al. as applied to claims 1, 22, and 23 above, and further in view of U. S. Patent No. 5,807,397 to Barreras.

Snell et al. disclose implanting in the patient a pulse generator (pulse generation and delivery circuit 60 shown on Fig. 1) having a power source (battery 70 shown on Fig. 1), and a

lead (lead system 30 shown on Fig. 1)) connected to the pulse generator and stimulating nervous tissue with electrical pulses generated by the pulse generator and communicated by the lead and determining the status and remaining life of the power source (see col. 7, lines 29-57).

Snell et al. in combination with Merritt et al. do not disclose controlling the pulse generator within preset limits by patient to adjust stimulation of nervous tissue.

Barreras teaches controlling the pulse generator within preset limits by patient to adjust stimulation of nervous tissue (see col. 18, lines 31-46 and col. 19, lines 18-23).

It would have been obvious to one having ordinary skill in the art at the time the invention was made to modify Snell et al. in combination with Merritt et al.'s method to include controlling the pulse generator within preset limits by patient to adjust stimulation of nervous tissue, as taught by Barreras, in order that stimulation parameters or control values in the implanted stimulator can be adjusted within preset limits to function effectively.

12. Claim 39 is rejected under 35 U.S.C. 103(a) as being unpatentable over Snell et al. in view of Merritt et al. and Canny et al. as applied to claims 1, 22, 25, 26, and 27 above, and further in view of U. S. Patent No. 5,807,397 to Barreras.

As to claim 39, Snell et al. disclose implanting in the patient a pulse generator (pulse generation and delivery circuit 60 shown on Fig. 1) having a power source (battery 70 shown on Fig. 1), and a lead (lead system 30 shown on Fig. 1) connected to the pulse generator and stimulating nervous tissue with electrical pulses generated by the pulse generator and communicated by the lead and determining the status and remaining life of the power source (see col. 7, lines 29-57)

Snell et al. in combination with Merritt et al. and Canny et al. do not disclose controlling the pulse generator within preset limits by patient to adjust stimulation of nervous tissue.

Barreras teaches controlling the pulse generator within preset limits by patient to adjust stimulation of nervous tissue (see col. 18, lines 31-46 and col. 19, lines 18-23).

It would have been obvious to one having ordinary skill in the art at the time the invention was made to modify Snell et al. in combination of Merritt et al. and Canny et al.'s method to include controlling the pulse generator within preset limits by patient to adjust stimulation of nervous tissue, as taught by Barreras, in order that stimulation parameters or control values in the implanted stimulator can be adjusted within preset limits to function effectively.

Allowable Subject Matter

13. Claims 17 and 20 are allowed.

U. S. Patent No. 5,344,431 to Merritt et al. is the reference closest to the claimed invention. Merritt et al. disclose a method of determining the current status and remaining life of a power source in an implantable neurological tissue stimulator comprising the steps of: assessing the power source voltage of the power source in an implantable neurological tissue stimulator; determining, based on the assessed power source voltage, where the power source is in its power source life cycle; and taking appropriate action in response to the determination of where the power source is in its power source life cycle. However, Merritt et al. do not teach calculating the remaining power source capacity by using a formula of the form: Remaining Battery Capacity = a constant + a constant multiplied by the power source voltage determined in

the step of assessing the power source voltage of the power source in an implantable neurological tissue stimulator; and including all of the other limitations in the respective independent claims.

Response to Arguments

14. Applicant's arguments with respect to claims 1-16, 18, 19, and 21-39 have been considered but are moot in view of the new ground(s) of rejection.

Contact Information

15. Any inquiry concerning this communication or earlier communications from the examiner should be directed to Carol S. Tsai whose telephone number is (703) 305-0851. The examiner can normally be reached on Monday-Friday from 7:30 AM to 4:00 PM. If attempts to reach the examiner by telephone are unsuccessful, the examiner's supervisor, Marc S. Hoff can be reached on (703) 308-1677. The fax number for TC 2800 is (703) 308-7382. Any inquiry of a general nature or relating to the status of this application or proceeding should be directed to the TC 2800 receptionist whose telephone number is (703) 308-1782.

In order to reduce pendency and avoid potential delays, Group 2800 is encouraging FAXing of responses to Office actions directly into the Group at (703) 308-7382. This practice may be used for filing papers not requiring a fee. It may also be used for filing papers which require a fee by applicants who authorize charges to a PTO deposit account. Please identify the

- Application/Control Number: 09/364,967
Art Unit: 2857

Page 13

examiner and art unit at the top of your cover sheet. Papers submitted via FAX into Group 2800 will be promptly forwarded to the examiner.

Carol S. Tsai

01/24/03


MARC S. HOFF
SUPERVISORY PATENT EXAMINER
TECHNOLOGY CENTER 2800